

Expert Advice

v2.0



Malvern Instruments Ltd - © Copyright 2008

Size measurements

Insufficient signal collected (5774 < 10000 k total photons).

Measurement duration not set to automatic.

Filter factor in research software not set to default (50%).

Cumulant fit error high.

Data quality too poor for cumulant analysis.

Sample too polydisperse for cumulant analysis.