

Size Distribution Report by Intensity

v2.1



Sample Details

Sample Name: Muestra 1 Promedio

SOP Name: mansettings.nano

General Notes: Average result created from record number(s): 1 2 3 4 5

File Name: 030418.dts

Dispersant Name: Water

Record Number: 6

Dispersant RI: 1,330

Material RI: 1,49

Viscosity (cP): 0,8872

Material Absorbtion: 0,001

Measurement Date and Time: martes, 03 de abril de 2018 1...

System

Temperature (°C): 25,0

Duration Used (s): 20

Count Rate (kcps): 141,0

Measurement Position (mm): 1,05

Cell Description: Disposable sizing cuvette

Attenuator: 8

Results

		Size (r.nm):	% Intensity	Width (r.nm):
Z-Average (r.nm): 8,762	Peak 1:	9,266	83,9	4,891
Pdl: 0,351	Peak 2:	259,2	13,6	223,7
Intercept: 0,800	Peak 3:	1783	2,5	574,2

Result quality : **Refer to quality report**

